

July 10, 2023

Jesse Bell
NB-Tri-Valley Central School District
34 Moore Hill Rd
Grahamsville, NY 12740

Project Location: Various Sampling Points
Project Number: Trivalley Central School District
Laboratory Work Order Number: 23F3202
PWSID# 5202956

Enclosed are results of analyses for samples received by the laboratory on June 23, 2023. If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Project Manager



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

NB-Tri-Valley Central School District
34 Moore Hill Rd
Grahamsville, NY 12740
ATTN: Jesse Bell

REPORT DATE: 7/10/2023

PURCHASE ORDER NUMBER: 23-00130

PROJECT NUMBER: Trivalley Central School District

ANALYTICAL SUMMARY

WORK ORDER NUMBER: 23F3202

The results of analyses performed on the following samples submitted to Pace Analytical Services, LLC - Newburgh are found in this report.

PROJECT LOCATION: Various Sampling Points

FIELD SAMPLE #	LAB ID:	MATRIX	SAMPLE DESCRIPTION	TEST	SUB LAB
Rm 267 Sink	23F3202-01	Drinking Water		EPA 200.8 Rev 5.4	
Rm 267 Fountain	23F3202-02	Drinking Water		EPA 200.8 Rev 5.4	
Rm 266 Sink	23F3202-03	Drinking Water		EPA 200.8 Rev 5.4	
Rm 266 Fountain	23F3202-04	Drinking Water		EPA 200.8 Rev 5.4	
Rm 265 Fountain	23F3202-05	Drinking Water		EPA 200.8 Rev 5.4	
Rm 265 Sink	23F3202-06	Drinking Water		EPA 200.8 Rev 5.4	
Rm 264 Fountain	23F3202-07	Drinking Water		EPA 200.8 Rev 5.4	
Rm 264 Sink	23F3202-08	Drinking Water		EPA 200.8 Rev 5.4	
W+R Fountain By 262	23F3202-09	Drinking Water		EPA 200.8 Rev 5.4	
Rm 263 Sink	23F3202-10	Drinking Water		EPA 200.8 Rev 5.4	
Rm 263 Fountain	23F3202-11	Drinking Water		EPA 200.8 Rev 5.4	
Rm 261 Sink	23F3202-12	Drinking Water		EPA 200.8 Rev 5.4	
Rm 261 Fountain	23F3202-13	Drinking Water		EPA 200.8 Rev 5.4	
Rm 260 Fountain	23F3202-14	Drinking Water		EPA 200.8 Rev 5.4	
Rm 260 Sink	23F3202-15	Drinking Water		EPA 200.8 Rev 5.4	
Rm 259 Sink	23F3202-16	Drinking Water		EPA 200.8 Rev 5.4	
Rm 259 Fountain	23F3202-17	Drinking Water		EPA 200.8 Rev 5.4	
Rm 258 Fountain	23F3202-18	Drinking Water		EPA 200.8 Rev 5.4	
Rm 258 Sink	23F3202-19	Drinking Water		EPA 200.8 Rev 5.4	
Rm 257 Fountain	23F3202-20	Drinking Water		EPA 200.8 Rev 5.4	
Rm 257 Sink	23F3202-21	Drinking Water		EPA 200.8 Rev 5.4	
ES Boys B.R. L Sink By 256	23F3202-22	Drinking Water		EPA 200.8 Rev 5.4	
ES Boys B.R. R Sink By 256	23F3202-23	Drinking Water		EPA 200.8 Rev 5.4	
ES Girls B.R. L Sink By 256	23F3202-24	Drinking Water		EPA 200.8 Rev 5.4	
ES Girls B.R. R Sink By 256	23F3202-25	Drinking Water		EPA 200.8 Rev 5.4	
ES Water Fountain Across 255	23F3202-26	Drinking Water		EPA 200.8 Rev 5.4	
Rm 256 Fountain	23F3202-27	Drinking Water		EPA 200.8 Rev 5.4	
Rm 256 Sink	23F3202-28	Drinking Water		EPA 200.8 Rev 5.4	
Rm 255 Fountain	23F3202-29	Drinking Water		EPA 200.8 Rev 5.4	
Rm 255 Sink	23F3202-30	Drinking Water		EPA 200.8 Rev 5.4	

CASE NARRATIVE SUMMARY

All reported results are within defined laboratory quality control objectives unless listed below or otherwise qualified in this report.

The results of analyses reported only relate to samples submitted to the Pace Analytical Services, LLC - Newburgh for testing.

I certify that the analyses listed above, unless specifically listed as subcontracted, if any, were performed under my direction according to the approved methodologies listed in this document, and that based upon my inquiry of those individuals immediately responsible for obtaining the information, the material contained in this report is, to the best of my knowledge and belief, accurate and complete.



Felicia Morgan-Nichols
Project Manager



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 267 Sink

Sampled: 6/23/2023 06:02

Sample ID: 23F3202-01

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	180	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:29	EMC
Lead	1.7	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 16:57	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 267 Fountain

Sampled: 6/23/2023 06:02

Sample ID: 23F3202-02

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	110	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:32	EMC
Lead	6.4	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:07	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 266 Sink

Sampled: 6/23/2023 06:05

Sample ID: 23F3202-03

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	150	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:34	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:09	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 266 Fountain

Sampled: 6/23/2023 06:05

Sample ID: 23F3202-04

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	210	10	1300	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:12	JC
Lead	2.2	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:12	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 265 Fountain

Sampled: 6/23/2023 06:09

Sample ID: 23F3202-05

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	240	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:36	EMC
Lead	6.3	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:14	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 265 Sink

Sampled: 6/23/2023 06:08

Sample ID: 23F3202-06

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	110	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:39	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:17	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 264 Fountain

Sampled: 6/23/2023 06:11

Sample ID: 23F3202-07

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	230	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:41	EMC
Lead	1.3	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:19	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 264 Sink

Sampled: 6/23/2023 06:11

Sample ID: 23F3202-08

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	210	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:44	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:27	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: W+R Fountain By 262

Sampled: 6/23/2023 06:16

Sample ID: 23F3202-09

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	29	10	1300	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:29	JC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:29	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 263 Sink

Sampled: 6/23/2023 06:17

Sample ID: 23F3202-10

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	280	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:51	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:31	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 263 Fountain

Sampled: 6/23/2023 06:17

Sample ID: 23F3202-11

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	140	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:53	EMC
Lead	3.6	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:34	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 261 Sink

Sampled: 6/23/2023 06:21

Sample ID: 23F3202-12

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	280	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:56	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:39	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 261 Fountain

Sampled: 6/23/2023 06:21

Sample ID: 23F3202-13

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	150	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 16:58	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:41	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 260 Fountain

Sampled: 6/23/2023 06:24

Sample ID: 23F3202-14

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	610	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/5/23	7/7/23 11:35	EMC
Lead	1.5	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:44	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 260 Sink

Sampled: 6/23/2023 06:24

Sample ID: 23F3202-15

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	140	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:03	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:46	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 259 Sink

Sampled: 6/23/2023 06:27

Sample ID: 23F3202-16

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	190	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:06	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:48	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 259 Fountain

Sampled: 6/23/2023 06:28

Sample ID: 23F3202-17

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	210	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:08	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:56	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 258 Fountain

Sampled: 6/23/2023 06:30

Sample ID: 23F3202-18

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	120	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:10	EMC
Lead	1.1	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 17:58	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 258 Sink

Sampled: 6/23/2023 06:30

Sample ID: 23F3202-19

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	480	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/5/23	7/7/23 11:38	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:01	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 257 Fountain

Sampled: 6/23/2023 06:34

Sample ID: 23F3202-20

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	500	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/5/23	7/7/23 11:40	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:03	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 257 Sink

Sampled: 6/23/2023 06:34

Sample ID: 23F3202-21

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	360	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:22	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:05	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: ES Boys B.R. L Sink By 256

Sampled: 6/23/2023 06:39

Sample ID: 23F3202-22

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	440	40	1300	µg/L	4		EPA 200.8 Rev 5.4	7/5/23	7/7/23 11:43	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:15	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: ES Boys B.R. R Sink By 256

Sampled: 6/23/2023 06:39

Sample ID: 23F3202-23

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	200	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:27	EMC
Lead	1.4	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:18	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: ES Girls B.R. L Sink By 256

Sampled: 6/23/2023 06:42

Sample ID: 23F3202-24

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	560	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/5/23	7/7/23 11:45	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:30	EMC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: ES Girls B.R. R Sink By 256

Sampled: 6/23/2023 06:43

Sample ID: 23F3202-25

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	210	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:32	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:27	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: ES Water Fountain Across 255

Sampled: 6/23/2023 06:44

Sample ID: 23F3202-26

Sample Matrix: Drinking Water

Metals Analyses (Total)										
Analyte	Results	MCL/SMCL		Units	Dilution	Flag/Qual	Method	Date	Date/Time	Analyst
		RL						Prepared	Analyzed	
Copper	210	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:35	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:30	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 256 Fountain

Sampled: 6/23/2023 06:43

Sample ID: 23F3202-27

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	110	100	1300	µg/L	10		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:37	EMC
Lead	1.1	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:32	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Sampled: 6/23/2023 06:43

Field Sample #: Rm 256 Sink

Sample ID: 23F3202-28

Sample Matrix: Drinking Water

Metals Analyses (Total)										
Analyte	Results	MCL/SMCL		Units	Dilution	Flag/Qual	Method	Date	Date/Time	Analyst
		RL						Prepared	Analyzed	
Copper	200	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:39	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:35	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 255 Fountain

Sampled: 6/23/2023 06:46

Sample ID: 23F3202-29

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	190	50	1300	µg/L	5		EPA 200.8 Rev 5.4	7/3/23	7/5/23 17:42	EMC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:37	JC



315 Fullerton Avenue * Newburgh, NY 12550 * TEL. (845) 562-0890

Project Location: Various Sampling Points

Sample Description:

Work Order: 23F3202

Date Received: 6/23/2023

Field Sample #: Rm 255 Sink

Sampled: 6/23/2023 06:46

Sample ID: 23F3202-30

Sample Matrix: Drinking Water

Metals Analyses (Total)

Analyte	Results	RL	MCL/SMCL	Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
Copper	160	10	1300	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:39	JC
Lead	ND	1.0	15	µg/L	1		EPA 200.8 Rev 5.4	6/29/23	6/30/23 18:39	JC

FLAG/QUALIFIER SUMMARY

*	QC result is outside of established limits.
†	Wide recovery limits established for difficult compound.
‡	Wide RPD limits established for difficult compound.
#	Data exceeded client recommended or regulatory level
ND	Not Detected
RL	Reporting Limit is at the level of quantitation (LOQ)
DL	Detection Limit is the lower limit of detection determined by the MDL study
MCL	Maximum Contaminant Level
	Percent recoveries and relative percent differences (RPDs) are determined by the software using values in the calculation which have not been rounded.
	No results have been blank subtracted unless specified in the case narrative section.

CERTIFICATIONS

Certified Analyses included in this Report

Analyte	Certifications
<i>EPA 200.8 Rev 5.4 in Drinking Water</i>	
Copper	NB-CT,NB-NJ,NB-NY
Lead	NB-CT,NB-NJ,NB-NY

Pace Analytical Services, LCC operates under the following certifications and accreditations:

Code	Description	Number	Expires
NB-CT	Connecticut Department of Public Health	PH-0823	09/30/2024
NB-NJ	New Jersey DEP	NY015 NELAP	06/30/2023
NB-NY	New York State Department of Health	10142 NELAP	04/1/2024

ANALYTICAL SERVICES

CHAIN OF CUSTODY

Lab Name	PAS - Newburgh	NYS DOH LAB # 10142	NIDEF LAB # NY105	CT DOPH# PH-0855
Lab Address	315 Fullerton Avenue, Newburgh, NY 12550	Phone (845) 562-0890		
Field Office Address	312 Titusville Rd, Poughkeepsie, NY 12603		Field Office Phone (845) 229-5536	
Field Office Address	35 Goshen Turnpike, Bloomingburg, NY 12721		Field Office Phone (845)733-1557	

23F3202

[illegible]



ANALYTICAL SERVICES

CHAIN OF CUSTODY

Lab Name PAS - Newburgh
Lab Address 315 Fullerton Avenue, Newburgh, NY 12550 Phone (845) 562-0890
Field Office Address 312 Titusville Rd, Poughkeepsie, NY 12603
Field Office Phone (845) 228-6536

NYS DOH LAB # 10142 NUDER LAB # NY105 CT DOPH# PH-0555
Field Office Address 38 Goshen Turnpike, Bloomingburg, NY 12721
Field Office Phone (845) 733-1557

REPORT# (Lab Use Only)

23F3202

PAGE 2 of 3

CLIENT NAME		PWS NUMBER		MATRIX TYPE		REQUIRED Containers												TURNAROUND TIME (Biz Days)				NON-TESTING CHARGES																					
CLIENT ADDRESS		5202956		34 MORE HILL RD GOSHENVILLE NY 12740		D (Drinking Water) or W (Waste Water) Indicate		Total # of Containers		40mL Vials HCl		Liter Amber HCl		250 Amber Sulfuric		Liter Amber Glass		250 mL Plastic Nitric Acid		250mL Plastic Sulfuric Acid		Liter Plastic		250mL Plastic		250mL Plastic NaOH		40mL Vials Sulfuric		40 mL Glass Plain		125 mL Sterile Na2S2O3		125mL Sterile		Other		NORMAL		P/U		SAMP	
CLIENT PHONE1		845-985-2896		Jesebelle		SOLID OR SEMISOLID		250mL Plastic Sulfuric		Liter Plastic		250mL Plastic		250mL Plastic NaOH		40mL Vials Sulfuric		40 mL Glass Plain		125 mL Sterile Na2S2O3		125mL Sterile		Other		RUSH (Y/N)		GRAB		COMP		RUSH (# Biz Days)		REPORTING									
EMAIL/TO SEND REPORT		Jesebelle		P.O. NUMBER/PROJECT NUMBER		Chlorine Residual		250mL Plastic Sulfuric		Liter Plastic		250mL Plastic		250mL Plastic NaOH		40mL Vials Sulfuric		40 mL Glass Plain		125 mL Sterile Na2S2O3		125mL Sterile		Other		RUSH (# Biz Days)		REPORTING															
PROJECT LOCATION		Jesebelle		0109				250mL Plastic Sulfuric		Liter Plastic		250mL Plastic		250mL Plastic NaOH		40mL Vials Sulfuric		40 mL Glass Plain		125 mL Sterile Na2S2O3		125mL Sterile		Other		RUSH (# Biz Days)		REPORTING															
SAMPLE DATE		TIME		SAMPLE IDENTIFICATION		COMPOSITE (C) OR GRAB (G) INDICATE		AQUEOUS (WATER)		D (Drinking Water) or W (Waste Water) Indicate		SOLID OR SEMISOLID		Chlorine Residual		NUMBER OF CONTAINERS SUBMITTED		Analysis Requested																									
6-23-23		6:12:11		Km 261 Fountain												1		dead/corpe																									
6:12:14		Km 260 Fountain														1																											
6:12:19		Km 259 Sink														1																											
6:12:27		Km 259 Sink														1																											
6:12:28		Km 259 Fountain														1																											
6:13:00		Km 258 Fountain														1																											
6:13:00		Km 258 Sink														1																											
6:13:14		Km 257 Fountain														1																											
6:13:34		Km 257 Sink														1																											
6:13:39		SS bags B.L.Sink By 256														1																											
6:13:39		SS bags B.L.Sink By 256														1																											
6:13:39		SS bags B.L.Sink By 256														1																											
6:14:2		SS bags B.L.Sink By 256														1																											

NOTES:		RECEIVED BY: (SIGNATURE)		DATE		TIME		RECEIVED BY: (SIGNATURE)		DATE		TIME	
SAMPLED BY: (SIGNATURE)		COMPANY		6-23-23		7:10		COMPANY		6-23-23		10:34	
RELINQUISHED BY: (SIGNATURE)		COMPANY		6-23-23		10:34		COMPANY		6-23-23		10:34	
RELINQUISHED BY: (SIGNATURE)		COMPANY		6-23-23		10:34		COMPANY		6-23-23		10:34	
RECEIVED FOR PAS-NEWBURGH BY: (SIGNATURE)		DATE		TIME		CUSTODY INTACT		YES		NO		COOLING TEMP	
1. ROSS PAGE		JUN 23 2023		15:45		YES		NO		COOLING TEMP		PAS-NEWBURGH REMARKS: ICE (Y/N) pH/Preservation Check	

Sample Condition Upon Receipt Form (SCUR)

Project # 23F3202
Client: Tri Valley CSD

Date and Initials of person:
Examining contents: JE
Label: _____
Deliver to location: Bloomington
pH: _____

Thermometer Used: NA

Date: 6/23/23 Time: 10:34 Initials: JE

State of Origin: NY

Cooler #1 Temp. °C _____ (Visual) _____ (Correction Factor) _____ (Actual)

☐ Samples on ice, cooling process has begun

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☒ Client ☐ Commercial ☐ Pace

☐ Other _____

Shipping Method: ☐ First Overnight ☐ Priority Overnight ☐ Standard Overnight ☐ Ground

☐ Other _____

Tracking # _____

Custody Seal on Cooler/Box Present: ☐ Yes ☐ No Seals intact: ☐ Yes ☐ No Ice: Wet Blue Melted None

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☐ None ☐ Other _____

Samples were collected by Pace employee: ☐ Yes ☐ No ☐ N/A

Comments:

Chain of Custody Present	☒ Yes ☐ No ☐ N/A	
Chain of Custody Filled Out	☒ Yes ☐ No ☐ N/A	
Relinquished Signature on CDC	☒ Yes ☐ No ☐ N/A	
Sampler Name and Signature on CDC	☒ Yes ☐ No ☐ N/A	
Samples Arrived within Hold Time	☒ Yes ☐ No ☐ N/A	
Rush TAJ requested on CDC	☐ Yes ☐ No ☒ N/A	
Sufficient Volume	☒ Yes ☐ No ☐ N/A	
Correct Containers Used	☒ Yes ☐ No ☐ N/A	
Containers Intact	☒ Yes ☐ No ☐ N/A	
Sample Labels match CDC (sample IDs & date/time of collection)	☒ Yes ☐ No ☐ N/A	
All containers needing acid/base preservation have been checked.	☒ Yes ☐ No ☐ N/A	Preservation Information: Preservative: _____ Lot #/Trace #: _____ Date: _____ Time: _____ Initials: _____
All Containers needing preservation are found to be in compliance with EPA recommendation: Exceptions: Vials, Microbiology, O&G, Metals	☒ Yes ☐ No ☐ N/A	
Headspace in VOA Vials? (>6mm):	☐ Yes ☐ No ☒ N/A	
Trip Blank Present	☐ Yes ☐ No ☒ N/A	

Additional Login Comments: _____

Client notification/ Resolution _____
Person Contacted: _____ Date/Time: _____
Comments/Resolution: _____